

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUL-0061  
**Location:** EVESHAM TOWNSHIP P.D.  
**Calibration File No.:** 01293 **Calib. Date:** 10/07/2013 **Calib. No.:** 00018  
**Certification File No.:** 01294 **Cert. Date:** 10/07/2013 **Cert. No.:** 00015  
**Linearity File No.:** 01295 **Lin. Date:** 10/07/2013 **Lin. No.:** 00015  
**Solution File No.:** 01313 **Soln. Date:** 12/07/2013 **Soln. No.:** 00147  
**Sequential File No.:** 01313 **File Date:** 12/07/2013

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDUNS3-0310  
**Control Solution %:** 0.100% **Expires:** 03/02/2014  
**Solution Control Lot:** 12C100 **Bottle No.:** 1076

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:32S		
Control 1 EC	0.101%	09:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33S		
Control 2 EC	0.099%	09:34S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:35S		
Control 3 EC	0.099%	09:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:36S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ARMSTRONG

First Name: JACK

MI: J

Signature: ofc. Armstrong #93

Badge No.: 93

Date: 12/07/2013

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUL-0061
Location:	EVESHAM TOWNSHIP P.D.	Calib. No.:	00018
Calibration File No.:	01293	Calib. Date:	10/07/2013
Certification File No.:	01294	Cert. Date:	10/07/2013
Linearity File No.:	01295	Lin. Date:	10/07/2013
Solution File No.:	01303	Soln. Date:	11/08/2013
Sequential File No.:	01303	File Date:	11/08/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUNS3-0310
Solution Control Lot:	12C100	Expires:	03/02/2014
		Bottle No.:	1100

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:30S		
Control 1 EC	0.101%	10:31S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:32S		
Control 2 EC	0.099%	10:32S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	10:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:33S		
Control 3 EC	0.100%	10:34S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: **MAGEE**

First Name: **THOMAS**

MI: **A**

Signature: Patrolman Thomas Magee #50

Badge No.: **50**

Date: **11/08/2013**

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUL-0061
Location:	EVESHAM TOWNSHIP P.D.		
Calibration File No.:	01293	Calib. Date:	10/07/2013
Certification File No.:	01294	Cert. Date:	10/07/2013
Linearity File No.:	01295	Lin. Date:	10/07/2013
Solution File No.:	01296	Soln. Date:	10/07/2013
Sequential File No.:	01296	File Date:	10/07/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDDC-0002
Solution Control Lot:	12C100	Expires:	03/02/2014
		Bottle No.:	1084

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:59D		
Control 1 EC	0.101%	14:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:01D		
Control 2 EC	0.101%	14:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		
Control 3 EC	0.100%	14:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDMAPI-0034 NYC

**Changed By:**

Last Name: GONCALVES                      First Name: MICHELLE                      MI: L

Signature: A. Sgr. Delos # 6040                      Badge No.: 6040                      Date: 10/07/2013

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUL-0061  
**Location:** EVESHAM TOWNSHIP P.D.  
**Calibration File No.:** 01230 **Calib. Date:** 04/29/2013 **Calib. No.:** 00017  
**Certification File No.:** 01231 **Cert. Date:** 04/29/2013 **Cert. No.:** 00014  
**Linearity File No.:** 01232 **Lin. Date:** 04/29/2013 **Lin. No.:** 00014  
**Solution File No.:** 01283 **Soln. Date:** 09/13/2013 **Soln. No.:** 00144  
**Sequential File No.:** 01283 **File Date:** 09/13/2013

**Calibrating Unit:** WET **Model No.:** CU-34 **Serial No.:** DDUN S3-0310  
**Control Solution %:** 0.100% **Expires:** 03/02/2014  
**Solution Control Lot:** 12C100 **Bottle No.:** 1078

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:07D		
Control 1 EC	0.101%	12:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:09D		
Control 2 EC	0.099%	12:09D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:10D		
Control 3 EC	0.099%	12:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: RITTER

First Name: RONALD

MI: C

Signature: \_\_\_\_\_



Badge No.: 63

Date: 09/13/2013

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUL-0061
Location:	EVESHAM TOWNSHIP P.D.		
Calibration File No.:	01230	Calib. Date:	04/29/2013
Certification File No.:	01231	Cert. Date:	04/29/2013
Linearity File No.:	01232	Lin. Date:	04/29/2013
Solution File No.:	01280	Soln. Date:	08/18/2013
Sequential File No.:	01280	File Date:	08/18/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0310
Solution Control Lot:	12C100	Expires:	03/02/2014
		Bottle No.:	1085

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:22D		
Control 1 EC	0.099%	11:22D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:23D		
Control 2 EC	0.098%	11:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:24D		
Control 3 EC	0.098%	11:25D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCAMBIA

First Name: MARC

MI: F

Signature:  \*96

Badge No.: 2296

Date: 08/18/2013

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUL-0061
Location:	EVESHAM TOWNSHIP P.D.		
Calibration File No.:	01230	Calib. Date:	04/29/2013
Certification File No.:	01231	Cert. Date:	04/29/2013
Linearity File No.:	01232	Lin. Date:	04/29/2013
Solution File No.:	01270	Soln. Date:	07/20/2013
Sequential File No.:	01270	File Date:	07/20/2013
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0310
Solution Control Lot:	12C100	Expires:	03/02/2014
		Bottle No.:	1090

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:31D		
Control 1 EC	0.100%	08:32D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32D		
Control 2 EC	0.100%	08:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:33D		
Control 3 EC	0.099%	08:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	08:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DOUGHERTY

First Name: ANDREW

MI: B

Signature: \_\_\_\_\_

Badge No.: 104

Date: 07/20/2013

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUL-0061
Location:	EVESHAM TOWNSHIP P.D.	Calib. No.:	00017
Calibration File No.:	01230	Cert. Date:	04/29/2013
Certification File No.:	01231	Lin. Date:	04/29/2013
Linearity File No.:	01232	Soln. Date:	06/21/2013
Solution File No.:	01263	File Date:	06/21/2013
Sequential File No.:	01263		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUN S3-0310
Solution Control Lot:	12C100	Expires:	03/02/2014
		Bottle No.:	1098

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:56D		
Control 1 EC	0.099%	02:56D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	02:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:57D		
Control 2 EC	0.098%	02:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	02:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:58D		
Control 3 EC	0.099%	02:59D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MAGEE

First Name: THOMAS

MI: A

Signature: Patrolman Thomas Magee #50

Badge No.: 50

Date: 06/21/2013

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUL-0061
<b>Location:</b>	EVESHAM TOWNSHIP P.D.		
<b>Calibration File No.:</b>	01230	<b>Calib. Date:</b>	04/29/2013
<b>Certification File No.:</b>	01231	<b>Cert. No.:</b>	00017
<b>Linearity File No.:</b>	01232	<b>Cert. Date:</b>	04/29/2013
<b>Solution File No.:</b>	01262	<b>Lin. No.:</b>	00014
<b>Sequential File No.:</b>	01262	<b>Soln. Date:</b>	06/21/2013
		<b>Soln. No.:</b>	00140
		<b>File Date:</b>	06/21/2013
<b>Calibrating Unit:</b>	WET	<b>Model No.:</b>	CU-34
<b>Control Solution %:</b>	0.100%	<b>Serial No.:</b>	DDUN S3-0310
<b>Solution Control Lot:</b>	12C100	<b>Expires:</b>	03/02/2014
		<b>Bottle No.:</b>	1091

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	01:30D		
Control 1 EC	0.098%	01:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	01:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:31D		
Control 2 EC	0.097%	01:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	01:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:32D		
Control 3 EC	0.097%	01:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	01:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MAGEE

First Name: THOMAS

MI: A

Signature: \_\_\_\_\_

*Patrolman Thomas Magee #50*

Badge No.: 50

Date: 06/21/2013



# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUL-0061  
Location: EVESHAM TOWNSHIP P.D.  
Calibration File No.: 01230 Calib. Date: 04/29/2013 Calib. No.: 00017  
Certification File No.: 01231 Cert. Date: 04/29/2013 Cert. No.: 00014  
Linearity File No.: 01232 Lin. Date: 04/29/2013 Lin. No.: 00014  
Solution File No.: 01258 Soln. Date: 06/17/2013 Soln. No.: 00139  
Sequential File No.: 01258 File Date: 06/17/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0310  
Control Solution %: 0.100% Expires: 03/02/2014  
Solution Control Lot: 12C100 Bottle No.: 1094

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:31D		
Control 1 EC	0.099%	09:31D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	09:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:32D		
Control 2 EC	0.098%	09:33D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33D		
Control 3 EC	0.097%	09:34D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:34D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:35D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: RITTER

First Name: RONALD

MI: C

Signature: \_\_\_\_\_



Badge No.: 63

Date: 06/17/2013

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARUL-0061  
Location: EVESHAM TOWNSHIP P.D.  
Calibration File No.: 01230 Calib. Date: 04/29/2013 Calib. No.: 00017  
Certification File No.: 01231 Cert. Date: 04/29/2013 Cert. No.: 00014  
Linearity File No.: 01232 Lin. Date: 04/29/2013 Lin. No.: 00014  
Solution File No.: 01245 Soln. Date: 05/24/2013 Soln. No.: 00138  
Sequential File No.: 01245 File Date: 05/24/2013

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUN S3-0310  
Control Solution %: 0.100% Expires: 03/02/2014  
Solution Control Lot: 12C100 Bottle No.: 1087

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:24D		
Control 1 EC	0.098%	23:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:26D		
Control 2 EC	0.098%	23:26D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:27D		
Control 3 EC	0.098%	23:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ARMSTRONG

First Name: JACK

MI: J

Signature: 

Badge No.: 93

Date: 05/24/2013

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARUL-0061
Location:	EVESHAM TOWNSHIP P.D.	Calib. No.:	00017
Calibration File No.:	01230	Cert. Date:	04/29/2013
Certification File No.:	01231	Lin. Date:	04/29/2013
Linearity File No.:	01232	Soln. Date:	04/29/2013
Solution File No.:	01233	File Date:	04/29/2013
Sequential File No.:	01233		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUN S3-0310
Control Solution %:	0.100%			Expires:	03/02/2014
Solution Control Lot:	12C100			Bottle No.:	1092

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:43D		
Control 1 EC	0.101%	15:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:45D		
Control 2 EC	0.100%	15:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46D		
Control 3 EC	0.099%	15:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number DDW JPD-109 (new)

**Changed By:**

Last Name: GONCALVES

First Name: MICHELLE

MI: L

Signature: 

Badge No.: 6040

Date: 04/29/2013



State of New Jersey  
 OFFICE OF THE ATTORNEY GENERAL  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
 POST OFFICE BOX 7068  
 WEST TRENTON, NJ 08628-0068  
 (609) 882-2000

CHRIS CHRISTIE  
 Governor

JEFFREY S. CHIESA  
 Attorney General

KIM GUADAGNO  
 Lt. Governor

COLONEL JOSEPH R. FUENTES  
 Superintendent

**CERTIFICATION OF ANALYSIS**  
**0.10 PERCENT BREATH ALCOHOL SIMULATOR SOLUTION**

ACCEPTANCE SPECIFICATIONS FOR BREATH ALCOHOL SIMULATOR SOLUTION: Ethyl alcohol concentration within, but not exceeding, the range of 0.117 to 0.125 grams per 100 milliliters of solution.

MANUFACTURER: Draeger Safety, Inc.

ANALYSIS DATE: 3/21/2012

BREATH ALCOHOL SIMULATOR SOLUTION LOT NUMBER: 12C100

Representative samples of the above-referenced Lot Number were tested by Gas Chromatography and found to have a mean ethyl alcohol concentration range of 0.1202 to 0.1210 grams per 100 milliliters of solution.

This lot of breath alcohol simulator solution may be utilized as a known traceable standard for the purpose of conducting periodic tests, pursuant to N.J.A.C. 13:51-3.4, of approved breath test instruments (N.J.A.C. 13:51-3.5) utilized by law enforcement agencies in this State. The manufacturer's expiration date for this lot of breath alcohol simulator solution is March 2, 2014.

As Forensic Scientist III for the Division of State Police, I hereby certify and attest that the tests and results documented in this Certificate of Analysis were performed at the Office of Forensic Sciences of the Division of State Police on properly functioning and calibrated instruments and equipment. All procedures utilized are accurate, objective, and performed on a routine basis by personnel within the Office of Forensic Sciences, in accordance with their professional duties and responsibilities.

Monica L. Trapontin  
 Forensic Scientist III  
 NJSP Office of Forensic Sciences

Sworn to and subscribed before me this 23rd day of March, 2012.

Notary

MARY ELIZABETH MCLAUGHLIN  
 NOTARY PUBLIC OF NEW JERSEY  
 MY COMMISSION EXPIRES 12-24-2013



"An Internationally Accredited Agency"

New Jersey Is An Equal Opportunity Employer  
 Printed on Recycled Paper and Recyclable

