

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00699	Calib. Date: 09/29/2015
Certification File No.:	00700	Calib. No.: 00023
Linearity File No.:	00701	Cert. Date: 09/29/2015
Solution File No.:	00705	Cert. No.: 00019
Sequential File No.:	00705	Lin. Date: 09/29/2015
		Lin. No.: 00020
		Soln. Date: 10/11/2015
		Soln. No.: 00169
		File Date: 10/11/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 0519

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:25D		
Control 1 EC	0.097%	20:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	20:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:26D		
Control 2 EC	0.097%	20:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:27D		
Control 3 EC	0.097%	20:28D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.097%	20:28D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LUTZ

First Name: RALPH

MI: C

Signature: _____

RALPH LUTZ 2732

Badge No.: 2732

Date: 10/11/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00664	Calib. Date:	03/31/2015
Certification File No.:	00665	Cert. Date:	03/31/2015
Linearity File No.:	00666	Lin. Date:	03/31/2015
Solution File No.:	00692	Soln. Date:	08/31/2015
Sequential File No.:	00692	File Date:	08/31/2015

Calib. No.:	00022
Cert. No.:	00018
Lin. No.:	00019
Soln. No.:	00167

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUJ S3-0116
Control Solution %:	0.100%			Expires:	06/13/2017
Solution Control Lot:	15F139			Bottle No.:	0035

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:41D		
Control 1 EC	0.097%	11:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:42D		
Control 2 EC	0.096%	11:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44D		
Control 3 EC	0.097%	11:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT First Name: STEPHEN MI: P

Signature: 

Badge No.: 2730
Date: 08/31/2015

Signature: 

Badge No.: 2730
Date: 08/31/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00664	Calib. Date: 03/31/2015
Certification File No.:	00665	Cert. No.: 00022
Linearity File No.:	00666	Cert. Date: 03/31/2015
Solution File No.:	00688	Lin. No.: 00019
Sequential File No.:	00688	Soln. Date: 08/11/2015
		Soln. No.: 00166
		File Date: 08/11/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 0518

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:18D		
Control 1 EC	0.096%	14:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	14:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:20D		
Control 2 EC	0.096%	14:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	14:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:21D		
Control 3 EC	0.096%	14:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	14:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:22D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____

Sgt. Stephen G. 2730

Badge No.: 2730

Date: 08/11/2015

Signature: _____

Sgt. Stephen G. 2730

Badge No.: 2730

Date: 08/11/2015

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00664 **Calib. Date:** 03/31/2015 **Calib. No.:** 00022
Certification File No.: 00665 **Cert. Date:** 03/31/2015 **Cert. No.:** 00018
Linearity File No.: 00666 **Lin. Date:** 03/31/2015 **Lin. No.:** 00019
Solution File No.: 00685 **Soln. Date:** 07/03/2015 **Soln. No.:** 00165
Sequential File No.: 00685 **File Date:** 07/03/2015

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUJ S3-0116
Control Solution %: 0.100% **Expires:** 11/12/2015
Solution Control Lot: 13K125 **Bottle No.:** 1323

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:35D		
Control 1 EC	0.099%	14:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:37D		
Control 2 EC	0.098%	14:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:38D		
Control 3 EC	0.099%	14:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____

Sen Stephens - 2730

Badge No.: 2730

Date: 07/03/2015

Signature: _____

Sen Stephens - 2730

Badge No.: 2730

Date: 07/03/2015

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00664	Calib. Date: 03/31/2015
Certification File No.:	00665	Calib. No.: 00022
Linearity File No.:	00666	Cert. Date: 03/31/2015
Solution File No.:	00684	Cert. No.: 00018
Sequential File No.:	00684	Lin. Date: 03/31/2015
		Lin. No.: 00019
		Soln. Date: 07/03/2015
		Soln. No.: 00164
		File Date: 07/03/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1317


Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:03D		
Control 1 EC	----%	11:04D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	----%	11:04D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	----%	----		
Control 2 EC	----%	----	N/A	
Control 2 IR	----%	----	N/A	
Ambient Air Blank	----%	----		
Control 3 EC	----%	----	N/A	
Control 3 IR	----%	----	N/A	
Ambient Air Blank	----%	----		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT	First Name: STEPHEN	MI: P
Signature: <u></u>	Badge No.: 2730	Date: 07/03/2015

Signature: <u></u>	Badge No.: 2730	Date: 07/03/2015
---	-----------------	------------------

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARNK-0025
Location: WESTAMPTON POLICE
Calibration File No.: 00664 Calib. Date: 03/31/2015 Calib. No.: 00022
Certification File No.: 00665 Cert. Date: 03/31/2015 Cert. No.: 00018
Linearity File No.: 00666 Lin. Date: 03/31/2015 Lin. No.: 00019
Solution File No.: 00680 Soln. Date: 06/01/2015 Soln. No.: 00163
Sequential File No.: 00680 File Date: 06/01/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUJ S3-0116
Control Solution %: 0.100% Expires: 11/12/2015
Solution Control Lot: 13K125 Bottle No.: 0801

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:13D		
Control 1 EC	0.098%	20:13D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:14D		
Control 2 EC	0.098%	20:15D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	20:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:15D		
Control 3 EC	0.098%	20:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P


Signature: _____



Badge No.: 2730

Date: 06/01/2015

Signature: _____



Badge No.: 2730

Date: 06/01/2015

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00664	Calib. Date:	03/31/2015
Certification File No.:	00665	Cert. Date:	03/31/2015
Linearity File No.:	00666	Lin. Date:	03/31/2015
Solution File No.:	00679	Soln. Date:	06/01/2015
Sequential File No.:	00679	File Date:	06/01/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUJ S3-0116
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	0804

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:58D		
Control 1 EC	0.094%	14:58D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.095%	14:58D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	14:59D		
Control 2 EC	0.094%	15:00D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.096%	15:00D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:00D		
Control 3 EC	0.095%	15:01D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.096%	15:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:02D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature:

Sgt. Stephen ENT 2730

Badge No.: 2730

Date: 06/01/2015

Signature:

Sgt. Stephen ENT 2730

Badge No.: 2730

Date: 06/01/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00664	Calib. Date:	03/31/2015
Certification File No.:	00665	Cert. Date:	03/31/2015
Linearity File No.:	00666	Lin. Date:	03/31/2015
Solution File No.:	00674	Soln. Date:	05/05/2015
Sequential File No.:	00674	File Date:	05/05/2015
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUJ S3-0116
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	1316

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:14D		
Control 1 EC	0.101%	09:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:15D		
Control 2 EC	0.100%	09:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:17D		
Control 3 EC	0.100%	09:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____

Sgt. Stephen P...

Badge No.: 2730

Date: 05/05/2015

Signature: _____

Sgt. Stephen P...

Badge No.: 2730

Date: 05/05/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00664	Calib. Date: 03/31/2015
Certification File No.:	00665	Calib. No.: 00022
Linearity File No.:	00666	Cert. Date: 03/31/2015
Solution File No.:	00673	Cert. No.: 00018
Sequential File No.:	00673	Lin. Date: 03/31/2015
		Lin. No.: 00019
		Soln. Date: 05/01/2015
		Soln. No.: 00160
		File Date: 05/01/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DIDUJ S3-0116
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 1311

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:21D		
Control 1 EC	----%	15:22D	22.3°C	SIMULATOR TEMP. ERROR
Control 1 IR	----%	15:22D	22.3°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	----%	--:--		
Control 2 EC	----%	--:--	N/A	
Control 2 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		
Control 3 EC	----%	--:--	N/A	
Control 3 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: Sgt. Stephen ENT

Badge No.: 2730

Date: 05/01/2015

Signature: Sgt. Stephen ENT

Badge No.: 2730

Date: 05/01/2015

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00627	Calib. Date: 10/06/2014
Certification File No.:	00628	Calib. No.: 00021
Linearity File No.:	00629	Cert. Date: 10/06/2014
Solution File No.:	00658	Cert. No.: 00017
Sequential File No.:	00658	Lin. Date: 10/06/2014
		Lin. No.: 00018
		Soln. Date: 03/13/2015
		Soln. No.: 00158
		File Date: 03/13/2015
Calibrating Unit:	WET	Model No.: CU-14
Control Solution %:	0.100%	Serial No.: DDUJ S3-0116
Solution Control Lot:	13K125	Expires: 11/12/2015
		Bottle No.: 0805

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:56S		
Control 1 EC	0.099%	07:57S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:58S		
Control 2 EC	0.099%	07:58S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:59S		
Control 3 EC	0.099%	08:00S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:00S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:01S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:


Last Name: FERGUSON

First Name: BRIAN

MI: C

Signature: 

Badgc No.: 2736
Date: 03/13/2015

Signature: 

Badgc No.: 2736
Date: 03/13/2015

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARNK-0025
Location:	WESTAMPTON POLICE	
Calibration File No.:	00627	Calib. Date: 10/06/2014
Certification File No.:	00628	Cert. No.: 00021
Linearity File No.:	00629	Cert. Date: 10/06/2014
Solution File No.:	00651	Lin. No.: 00017
Sequential File No.:	00651	Lin. Date: 10/06/2014
		Lin. No.: 00018
		Soln. Date: 01/14/2015
		Soln. No.: 00156
		File Date: 01/14/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUJ S3-0116
Control Solution %:	0.100%		Expires: 11/12/2015
Solution Control Lot:	13K125		Bottle No.: 1313

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:29S		
Control 1 EC	0.100%	14:30S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:30S		
Control 2 EC	0.100%	14:31S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:31S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32S		
Control 3 EC	0.100%	14:32S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:32S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDSF P2 - 0433 MG

Changed By:

Last Name: GONCALVES First Name: MICHELLE MI: L

Signature: Sgt. Michelle Goncalves #6040 Badge No.: 6040
Date: 01/14/2015

Signature: Sgt. Michelle Goncalves #6040 Badge No.: 6040
Date: 01/14/2015



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 24th 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST[®] CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDUS 53-0116

Certification Date

1-5-15

Technician

BC

Re-Certification Due Date

1-5-16



Dräger

ALCOTEST[®] 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest[®] 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST). The manufacturer recommends accuracy verification of the Temperature Probe within 12 months of the certification date below, or sooner, according to your State Specification. For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest[®] 7110.

Serial Number Temp. Probe

DDSE P2 0433

Certification date:

1-5-15

Next Certification due:

1-5-16

Probe Value

106

Draeger Safety Diagnostics, Inc.
Technical Service Department

BC

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARNK-0025
Location:	WESTAMPTON POLICE		
Calibration File No.:	00627	Calib. Date:	10/06/2014
Certification File No.:	00628	Cert. Date:	10/06/2014
Linearity File No.:	00629	Lin. Date:	10/06/2014
Solution File No.:	00649	Soln. Date:	01/07/2015
Sequential File No.:	00649	File Date:	01/07/2015
Calibrating Unit: WET		Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUI S3-0117
Solution Control Lot:	13K125	Expires:	11/12/2015
		Bottle No.:	0802

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:33S		
Control 1 EC	0.101%	12:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:35S		
Control 2 EC	0.101%	12:35S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:36S		
Control 3 EC	0.101%	12:37S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:37S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:37S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ENT

First Name: STEPHEN

MI: P

Signature: _____

San. Stephen P... 2730

Badge No.: 2730

Date: 01/07/2015

Signature: _____

San. Stephen P... 2730

Badge No.: 2730

Date: 01/07/2015