

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00085 Calib. Date: 10/31/2006 Calib. No.: 00003
Certification File No.: 00086 Cert. Date: 10/31/2006 Cert. No.: 00002
Linearity File No.: 00087 Lin. Date: 10/31/2006 Lin. No.: 00002
Solution File No.: 00098 Soln. Date: 12/19/2006 Soln. No.: 00018
Sequential File No.: 00098 File Date: 12/19/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 04/14/2007
Solution Control Lot: 05D017 Bottle No.: 0219

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:59S		
Control 1 EC	0.098%	08:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	08:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00S		
Control 2 EC	0.097%	09:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01S		
Control 3 EC	0.096%	09:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 12/19/2006

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00085 Calib. Date: 10/31/2006 Calib. No.: 00003
Certification File No.: 00086 Cert. Date: 10/31/2006 Cert. No.: 00002
Linearity File No.: 00087 Lin. Date: 10/31/2006 Lin. No.: 00002
Solution File No.: 00094 Soln. Date: 11/25/2006 Soln. No.: 00017
Sequential File No.: 00094 File Date: 11/25/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 04/14/2007
Solution Control Lot: 05D017 Bottle No.: 0970

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:44S		
Control 1 EC	0.097%	09:45S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	09:45S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45S		
Control 2 EC	0.096%	09:46S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:46S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:47S		
Control 3 EC	0.096%	09:47S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	09:47S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: 

#

Badge No.: 102

Date: 11/25/2006

102

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00085	Calib. Date:	10/31/2006
Certification File No.:	00086	Cert. Date:	10/31/2006
Linearity File No.:	00087	Lin. Date:	10/31/2006
Solution File No.:	00088	Soln. Date:	10/31/2006
Sequential File No.:	00088	File Date:	10/31/2006
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWE S3-0188
Solution Control Lot:	05D017	Expires:	04/14/2007
		Bottle No.:	0974

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:04S		
Control 1 EC	0.099%	11:05S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:05S		
Control 2 EC	0.097%	11:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	11:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:07S		
Control 3 EC	0.096%	11:07S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	11:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUNP2. 389 WML

Changed By:

Last Name: CROSS	First Name: WILLIAM	MI: M
Signature: <u>TPR. I WML M. Cross #5342</u>	Badge No.: 5342	Date: 10/31/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00006	Calib. Date: 11/03/2005
Certification File No.:	00007	Cert. No.: 00002
Linearity File No.:	00008	Cert. Date: 11/03/2005
Solution File No.:	00081	Lin. No.: 00001
Sequential File No.:	00081	Lin. Date: 11/03/2005
		Soln. No.: 00015
		Soln. Date: 10/07/2006
		File Date: 10/07/2006

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0188
Control Solution %:	0.100%		Expires: 04/14/2007
Solution Control Lot:	05D017		Bottle No.: 0224

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:38D		
Control 1 EC	0.100%	15:38D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:39D		
Control 2 EC	0.098%	15:39D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	15:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:40D		
Control 3 EC	0.099%	15:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature:



Badge No.: 102

Date: 10/07/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00006 Calib. Date: 11/03/2005 Calib. No.: 00002
Certification File No.: 00007 Cert. Date: 11/03/2005 Cert. No.: 00001
Linearity File No.: 00008 Lin. Date: 11/03/2005 Lin. No.: 00001
Solution File No.: 00075 Soln. Date: 09/07/2006 Soln. No.: 00014
Sequential File No.: 00075 File Date: 09/07/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0188
Control Solution %: 0.100% Expires: 04/14/2007
Solution Control Lot: 05D017 Bottle No.: 0214

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	10:25D		
Control 1 EC	0.099%	10:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	10:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:26D		
Control 2 EC	0.098%	10:27D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	10:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28D		
Control 3 EC	0.098%	10:28D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo # 102

Badge No.: 102

Date: 09/07/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00006	Calib. Date: 11/03/2005
Certification File No.:	00007	Cert. No.: 00002
Linearity File No.:	00008	Cert. Date: 11/03/2005
Solution File No.:	00071	Lin. No.: 00001
Sequential File No.:	00071	Lin. Date: 11/03/2005
		Soln. No.: 00013
		Soln. Date: 08/11/2006
		File Date: 08/11/2006

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0188
Control Solution %:	0.100%			Expires:	04/14/2007
Solution Control Lot:	05D017			Bottle No.:	0966

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:59D		
Control 1 EC	0.100%	09:00D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00D		
Control 2 EC	0.099%	09:01D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:02D		
Control 3 EC	0.099%	09:02D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:02D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:03D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: _____

Sgt. Raymond Norfo #102

Badge No.: 102

Date: 08/11/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00006	Calib. Date:	11/03/2005
Certification File No.:	00007	Cert. Date:	11/03/2005
Linearity File No.:	00008	Lin. Date:	11/03/2005
Solution File No.:	00067	Soln. Date:	07/18/2006
Sequential File No.:	00067	File Date:	07/18/2006

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0188
Control Solution %:	0.100%			Expires:	04/14/2007
Solution Control Lot:	05D017			Bottle No.:	0160

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:36D		
Control 1 EC	0.101%	07:36D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:37D		
Control 2 EC	0.100%	07:37D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:38D		
Control 3 EC	0.100%	07:39D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:39D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature:

Sgt. Raymond Norfo # 102

Badge No.: 102

Date: 07/18/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWC-0060
Location:	WILLINGBORO TWP POLICE		
Calibration File No.:	00006	Calib. Date:	11/03/2005
Certification File No.:	00007	Cert. Date:	11/03/2005
Linearity File No.:	00008	Lin. Date:	11/03/2005
Solution File No.:	00061	Soln. Date:	06/20/2006
Sequential File No.:	00061	File Date:	06/20/2006

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWE S3-0188
Control Solution %:	0.100%			Expires:	04/14/2007
Solution Control Lot:	05D017			Bottle No.:	0204

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:51D		
Control 1 EC	0.100%	08:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:52D		
Control 2 EC	0.100%	08:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	08:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:54D		
Control 3 EC	0.099%	08:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	08:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NQRFO

First Name: RAYMOND

MI: A

Signature:

Sgt. Raymond Norfo #102

Badge No.: 102

Date: 06/20/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00006	Calib. Date: 11/03/2005
Certification File No.:	00007	Calib. No.: 00002
Linearity File No.:	00008	Cert. Date: 11/03/2005
Solution File No.:	00055	Cert. No.: 00001
Sequential File No.:	00055	Lin. Date: 11/03/2005
		Lin. No.: 00001
		Soln. Date: 05/24/2006
		Soln. No.: 00010
		File Date: 05/24/2006

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0188
Control Solution %:	0.100%		Expires: 04/14/2007
Solution Control Lot:	05D017		Bottle No.: 0154

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:04D		
Control 1 EC	0.099%	10:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05D		
Control 2 EC	0.099%	10:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06D		
Control 3 EC	0.099%	10:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

TEMPERATURE PROBE SERIAL NUMBER: DDUNP2-389 WMR

Changed By:

Last Name: CROSS

First Name: WILLIAM

MI: M

Signature: TPR. I WILLIAM M. CROSS #5342

Badge No.: 5342

Date: 05/24/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00006 Calib. Date: 11/03/2005 Calib. No.: 00002
Certification File No.: 00007 Cert. Date: 11/03/2005 Cert. No.: 00001
Linearity File No.: 00008 Lin. Date: 11/03/2005 Lin. No.: 00001
Solution File No.: 00051 Soln. Date: 04/25/2006 Soln. No.: 00009
Sequential File No.: 00051 File Date: 04/25/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 04/14/2007
Solution Control Lot: 05D017 Bottle No.: 0975

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:43D		
Control 1 EC	0.100%	07:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	07:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45D		
Control 2 EC	0.099%	07:45D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46D		
Control 3 EC	0.099%	07:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:47D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 04/25/2006

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00006 Calib. Date: 11/03/2005 Calib. No.: 00002
Certification File No.: 00007 Cert. Date: 11/03/2005 Cert. No.: 00001
Linearity File No.: 00008 Lin. Date: 11/03/2005 Lin. No.: 00001
Solution File No.: 00048 Soln. Date: 03/28/2006 Soln. No.: 00008
Sequential File No.: 00048 File Date: 03/28/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 04/14/2007
Solution Control Lot: 05D017 Bottle No.: 0153

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:45S		
Control 1 EC	0.100%	07:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46S		
Control 2 EC	0.100%	07:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	07:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:48S		
Control 3 EC	0.099%	07:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	07:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 03/28/2006

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00006	Calib. Date: 11/03/2005
Certification File No.:	00007	Cert. No.: 00002
Linearity File No.:	00008	Cert. Date: 11/03/2005
Solution File No.:	00039	Lin. No.: 00001
Sequential File No.:	00039	Lin. Date: 11/03/2005
		Soln. No.: 00007
		Soln. Date: 02/27/2006
		File Date: 02/27/2006

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0189
Control Solution %:	0.100%		Expires: 04/14/2007
Solution Control Lot:	05D017		Bottle No.: 0152

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:43S		
Control 1 EC	0.098%	07:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45S		
Control 2 EC	0.099%	07:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:46S		
Control 3 EC	0.098%	07:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
 Last Name: NORFO First Name: RAYMOND MI: A
 Signature:  # 102
 Badge No.: 102
 Date: 02/27/2006

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWC-0060
Location: WILLINGBORO TWP POLICE
Calibration File No.: 00006 Calib. Date: 11/03/2005 Calib. No.: 00002
Certification File No.: 00007 Cert. Date: 11/03/2005 Cert. No.: 00001
Linearity File No.: 00008 Lin. Date: 11/03/2005 Lin. No.: 00001
Solution File No.: 00026 Soln. Date: 01/31/2006 Soln. No.: 00006
Sequential File No.: 00026 File Date: 01/31/2006

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWE S3-0189
Control Solution %: 0.100% Expires: 04/14/2007
Solution Control Lot: 05D017 Bottle No.: 0223

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:59S		
Control 1 EC	0.100%	08:59S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:59S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:00S		
Control 2 EC	0.098%	09:01S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:01S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:01S		
Control 3 EC	0.099%	09:02S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature: Sgt. Raymond Norfo #102

Badge No.: 102

Date: 01/31/2006

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWC-0060
Location:	WILLINGBORO TWP POLICE	
Calibration File No.:	00006	Calib. Date: 11/03/2005
Certification File No.:	00007	Calib. No.: 00002
Linearity File No.:	00008	Cert. Date: 11/03/2005
Solution File No.:	00023	Cert. No.: 00001
Sequential File No.:	00023	Lin. Date: 11/03/2005
		Lin. No.: 00001
		Soln. Date: 01/03/2006
		Soln. No.: 00005
		File Date: 01/03/2006

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWE S3-0189
Control Solution %:	0.100%		Expires: 04/14/2007
Solution Control Lot:	05D017		Bottle No.: 0151

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:06S		
Control 1 EC	0.100%	00:07S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:08S		
Control 2 EC	0.100%	00:08S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:08S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:09S		
Control 3 EC	0.100%	00:10S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:10S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: NORFO

First Name: RAYMOND

MI: A

Signature:



Badge No.: 102

Date: 01/03/2006